

Notice of Allowability

Application No.

10/063,181

Examiner

Courtney Thomas

Applicant(s)

PETRICK ET AL.

Art Unit

2882

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 5/25/04.
2. ☒ The allowed claim(s) is/are 1-28.
3. ☒ The drawings filed on 21 March 2003 are accepted by the Examiner.
4. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

5. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 6. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
7. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

- | | |
|---|--|
| 1. <input type="checkbox"/> Notice of References Cited (PTO-892) | 5. <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 2. <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 6. <input type="checkbox"/> Interview Summary (PTO-413),
Paper No./Mail Date _____. |
| 3. <input type="checkbox"/> Information Disclosure Statements (PTO-1449 or PTO/SB/08),
Paper No./Mail Date _____ | 7. <input type="checkbox"/> Examiner's Amendment/Comment |
| 4. <input type="checkbox"/> Examiner's Comment Regarding Requirement for Deposit
of Biological Material | 8. <input checked="" type="checkbox"/> Examiner's Statement of Reasons for Allowance |
| | 9. <input type="checkbox"/> Other _____. |

DETAILED ACTION

Allowable Subject Matter

1. Claims 1-28 are allowed.
2. The following is an examiner's statement of reasons for allowance:
3. **As per claim 1 and dependent claims 2-9**, the examiner found no reference in the prior art that disclosed or made obvious a method to minimize signal conversion time for a solid state detector panel in a diagnostic X-ray system generating a plurality of images at varying frame rates, the method comprising the steps of a) measuring a set of induced signal offsets, whether positive or negative, caused by time varying charge retention associated with a detector panel during a phantom time segment prior to normal signal read-out of the detector panel for a current image frame and b) generating a set of adjustment values in response to the set of induced signal offsets as recited in independent claim 1.
4. **As per claim 10 and dependent claims 11-20**, the examiner found no reference in the prior art that disclosed or made obvious an apparatus to minimize signal conversion time for a diagnostic X-ray system generating a plurality of image frames at varying frame rates, the apparatus comprising read out electronics to read a current row of an array to be read, the read-out electronics being connected to columns of the array and being responsive to charge build-up to generate a set of normalized detector signals in response to phantom measurements such that the set of normalized detector signal is adjusted for offsets in signal strength, whether positive or negative, caused by temporal row-to-row variations in charge retention in the array as recited in independent claim 10.

Art Unit: 2882

5. As per claim 21 and dependent claims 22-28, the examiner found no reference in the prior art that disclosed or made obvious a diagnostic X-ray system comprising a) a solid state detector module responsive to X-ray signals and generating a plurality of charges representative of charge retention in response to phantom measurements and X-ray intensity and b) an image processing module responsive to the plurality of charges and generating a plurality of normalized detector signals for a current image frame, the normalized detector signals being dynamically adjusted for variations in charge retention, whether positive or negative, as frame rate changes as recited in independent claim 21.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Courtney Thomas whose telephone number is (571) 272-2496. The examiner can normally be reached on M - F (9 am - 5 pm).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Ed Glick can be reached on (571) 272 2490. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 2882

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

CT

Courtney Thomas



DAVID V. BRUCE
PRIMARY EXAMINER